



OHS ALERTI

Type of Incident	Explosion in an Autoclave and subsequent fire
Date Occurred	6 th October 2006
WHAT HAPPENED	An explosion occurred in the autoclave of the biological waste contractor.
	The biological waste contractor picked up biological waste from UNSW on 6 th October and took it to the waste processing centre at Unanderra.
	While decontaminating the biological waste an explosion occurred inside the autoclave. Upon removal of the waste from the autoclave it burst into flames.
	Early investigation indicates that glass vials with rubber diaphragms (GC or HPLC vial) containing chemical waste (probably a flammable liquid) were placed in the Biowaste bin. See attached Image 1 on page 2.
	The incident is currently being investigated by WorkCover and the EPA.
	The biological waste contractor is identifying the chemical found in some of the glass vials to try and determine its origin.
WERE THERE ANY INJURIES	None reported at this time
Most Likely Cause	Incorrect disposal of chemical waste in the Biowaste bin.
Corrective Action	 All areas that generate biological waste are to review their waste disposal procedures to ensure waste segregation. Ensure all staff and students are trained in biological waste and chemical waste disposal requirements. UNSW to conduct random biological waste inspections.

Image 1 - Some of the vials suspected of causing the incident.

